



# Size and Illumination Matters: Local Magnetic Actuation and Fluorescence Imaging for Microrobotics

Juan J. Huaroto<sup>1\*</sup> and Sarthak Misra<sup>1,2</sup>

**Abstract** | Combining local magnetic actuation with fluorescence imaging modalities promises to introduce significant advances in microrobotic-guided procedures. This review presents the advantages and challenges of this approach, emphasizing the need for careful design considerations to optimize performance and compatibility. Traditional microrobotic actuation systems rely on bulky electromagnets, which are unsuitable for clinical use due to high power requirements and limited operational workspace. In contrast, miniaturized electromagnets can be integrated into surgical instruments, offering low power consumption and high actuation forces at the target site. Fluorescence imaging modalities have been explored in microrobotics, showcasing spatiotemporal resolution and the capability to provide information from biological entities. However, limitations, such as shallow penetration depth and out-of-focus fluorescence, have motivated the development of advanced techniques such as two-photon microscopy. The potential of two-photon microscopy to overcome these limitations is highlighted, with supporting evidence from previous studies on rat tissue samples. Current challenges in optical penetration depth, temporal resolution, and field of view are also addressed in this review. While integrating miniaturized electromagnets with fluorescence imaging modalities holds the potential for microrobotic-guided procedures, ongoing research and technological advancements are essential to translating this approach into clinical practice.

**Keywords:** *Microrobotics, Surgical instruments, Miniaturized electromagnets, Fluorescence imaging, Two-photon microscopy.*

## 1 Introduction

In recent years, the field of microrobotics has witnessed significant advancements driven by the demand for precise manipulation and imaging at the microscale<sup>1-5</sup>. Microrobots or micro-agents serve as essential tools for navigating through intricate media, facilitating applications across various domains, including targeted drug delivery<sup>6</sup>, cell manipulation<sup>7</sup>, and minimally invasive surgery<sup>8</sup>. With the ongoing progress in microrobotic fabrication and functionalization, efficient actuation methods and imaging

modalities are crucial to unlocking novel applications and expanding the application domain toward clinically relevant scenarios.

Micro-agents can be engineered to respond to external stimuli such as magnetic fields<sup>9-11</sup>, acoustic waves<sup>12-14</sup>, and light<sup>15-17</sup>. In particular, magnetic actuation has emerged as a promising approach for propelling and maneuvering micro-agents in a contactless fashion<sup>18</sup>. Due to their transparency and biocompatibility, magnetic actuation has been utilized for micro-agent navigation in complex environments, achieving

<sup>1</sup> Surgical Robotics Laboratory, Department of Biomechanical Engineering, University of Twente, 7522 NB, Enschede, The Netherlands.

<sup>2</sup> Surgical Robotics Laboratory, Department of Biomaterials and Biomedical Technology, University Medical Centre Groningen and University of Groningen, 9713 GZ, Groningen, The Netherlands.

\*j.j.huarotosevilla@utwente.nl

groundbreaking applications under *ex vivo* and *in vivo* conditions<sup>19</sup>. By leveraging magnetic fields and gradients, magnetic actuation systems allow precise control without needing onboard power sources or complex mechanical components. In the present-day context, several magnetic actuation platforms have emerged following the use of bulky structures assembled to permanent magnets, electromagnets, and robotic platforms<sup>20–22</sup>. Motivated by the need to reduce the size and power consumption of magnetic systems, prior research has introduced the concept of local actuation via miniaturized magnets<sup>23–29</sup>. Miniaturized magnets provide direct access to target regions due to their size and potential integration into surgical instruments for in situ actuation of micro-agents.

In parallel with advancements in magnetic actuation, visualizing and tracking micro-agents in real-time is essential for understanding their dynamics and validating their interaction with biological entities<sup>30–33</sup>. Various imaging modalities have been proposed and adapted to visualize micro-agents to optimize spatiotemporal resolution, penetration depth, and the capability to obtain molecular information from living cells<sup>34–36</sup>. Among non-ionizing standard modalities such as magnetic resonance, photoacoustics, and ultrasound, fluorescence-based imaging techniques provide a powerful means of monitoring micro-agents with a relatively high spatiotemporal resolution but limited penetration depth<sup>37,38</sup>. Fluorescence imaging is an appealing modality for visualizing micro-agents alongside biological entities, permitting the study of the response to external stimuli with remarkable sensitivity<sup>7,39,40</sup>. Advanced fluorescence modalities such as two-photon microscopy are available in the literature to overcome the limited penetration depth of traditional fluorescence imaging<sup>41</sup>. However, additional efforts are required to bridge the gap between two-photon microscopy and microrobotics<sup>42,43</sup>. Other techniques, using higher-order photon absorption (e.g., three-photon microscopy)<sup>44</sup> and actively adjusting phase and amplitude of light wavefront (i.e., wavefront shaping)<sup>45</sup>, are promising for breaking the scattering limits of light in biological media. Nevertheless, the literature has not explored their integration into the microrobotics field.

The convergence of local magnetic actuation and fluorescence imaging in microrobotics presents an opportunity to unlock clinical advancements. The unique capabilities of fluorescence

imaging can be used for precise micro-agent tracking while offering insights into biological environments. Moreover, miniaturized magnets facilitate their integration into minimally invasive surgical instruments, enabling in situ micro-agent control. Merging both technologies represents an advancement in microrobotics, accelerating progress in research and clinical applications.

### 1.1 Scope and Outline


This review presents the current progress on magnetic actuation and fluorescence-based imaging to control and monitor micro-agents. Throughout the review, we present the types of magnetic actuation systems, emphasizing miniaturized electromagnets for local actuation of micro-agents. We discuss the need for advanced imaging modalities to achieve micro-agent in-tissue control. We highlight the principles of fluorescence imaging, its application to the microrobotics field, and current challenges in minimally invasive surgery. By combining local magnetic actuation and fluorescence imaging, we seek to address fundamental challenges in microrobotics, including precise manipulation in confined spaces and tracking within biological environments.

The remainder part of the review is divided into four sections. First, we summarize the current magnetic actuation systems, introduce the concept of local actuation, and present the equations governing microrobotic actuation using miniaturized electromagnets integrated into two types of minimally invasive surgical instruments. Second, we introduce fluorescence imaging for microrobotics, discuss the limitations of traditional modalities, and propose advanced techniques, such as two-photon microscopy, to address these shortcomings. Third, we explore the integration of miniaturized electromagnets and two-photon microscopy to enable fluorescence-guided micro-agent manipulation. Finally, we summarize and discuss the concepts presented throughout the review.

## 2 Electromagnetic Actuation in Microrobotics

### 2.1 Background

Magnetic actuation has been studied to enable contactless manipulation of micro/millirobotic structures and advanced catheters to access hard-to-reach regions of the human body<sup>9,21</sup>. In particular, the apparatus used in microrobotics has constantly evolved to address



**Figure 1:** Electromagnetic systems located outside the human body: Traditional arrangements. **A** Electromagnets assembled in a fixed platform. **B** Electromagnets assembled at the end effector of a robotic platform.

application-oriented requirements<sup>20,22,46,47</sup>. Permanent magnets and electromagnets are standard devices used in magnetic actuation systems to generate magnetic fields and gradients. Permanent magnets allow for the precise generation of magnetic fields at a defined distance from the magnet. However, the generation of time-varying magnetic fields is limited, necessitating robotic platforms to position or spin the permanent magnet with respect to a target region or workspace<sup>48–53</sup>. Unlike permanent magnets, electromagnets or electromagnetic coils can generate magnetic fields and gradients by powering them with electrical currents. In electromagnetic actuation systems, electromagnets can be arranged in stationary mechanical structures<sup>54–59</sup> (Fig. 1A) and robotic structures<sup>59–62</sup> to span a desired workspace (Fig. 1B).

In clinically relevant scenarios, electromagnetic systems are typically designed to be positioned externally, enabling non-invasive actuation of micro-agents inside the body. Traditional electromagnetic systems have successfully demonstrated medical applications using *in vivo* animal samples or human cadavers<sup>63–66</sup>. However, these systems are accompanied by various challenges that must be addressed to advance their translation to clinical settings. The magnetic field required to propel micro-agents within vessels must overcome physiological fluids to reach a desired region<sup>67</sup>. In practice, addressing the generation of relatively high magnetic fields and gradients often involves increasing the electromagnet size, adjusting electrical currents, or adding additional electromagnets to the actuation system<sup>68</sup>. However, the power consumption and space limitations of surgical rooms slow down the standardization of electromagnetic systems for human trials.

## 2.2 Miniaturized Systems

A different approach to addressing the challenges of traditional electromagnetic systems for micro-robotics involves the integration of miniaturized electromagnets into surgical instruments such as endoscopes<sup>25</sup> and laparoscopy probes<sup>26</sup> (Fig. 2). This integration offers the advantage of minimal invasiveness, making it a promising alternative in microsurgery where precise manipulation is required<sup>6</sup>. Compared with traditional electromagnetic systems, miniaturized electromagnets can significantly reduce power consumption (by three orders of magnitude) and provide direct access to biological workspaces while maintaining biologically compatible temperatures. Moreover, the magnetic field gradients near miniaturized electromagnets are comparable to the values registered at the workspace center of electromagnetic systems (Table 1).

Integrating miniaturized electromagnets into surgical instruments can revolutionize surgical procedures and introduce new treatments using magnetic micro-agents<sup>25,26</sup>. Previous studies have proposed such integration to provide basic functionalities, such as gripping small objects<sup>27,29</sup>. However, local actuation of micro-agents using miniaturized electromagnets presents challenges, requiring careful consideration of electric and thermal insulation. Additionally, unique strategies for the deployment and retrieval of miniaturized electromagnets need to be developed and tailored to surgical instruments.

For endoscopes, miniaturized electromagnets can be delivered through an internal channel and positioned around a target region using a deployment mechanism (Fig. 2A). This mechanism is deactivated when micro-agents reach a target region, permitting retrieval through the

**Table 1.** Magnetic field ( $\|\mathbf{B}\|$ ) and gradient-to-field ratio ( $\frac{\|\nabla\mathbf{B}\|}{\|\mathbf{B}\|}$ ) of external and local electromagnetic actuation systems. Table adapted from<sup>28</sup>

Configuration	Actuation system	Moving electromagnets	$\ \mathbf{B}\ $ [mT]	$\frac{\ \nabla\mathbf{B}\ }{\ \mathbf{B}\ }$ [m <sup>-1</sup> ]
External	Open Configuration <sup>57</sup>	No	15	35
	Antiprism <sup>55</sup>	No	30	40
	OctoMag <sup>54</sup>	No	30	25
	Toroidal <sup>69</sup>	No	50	5
	Maxwell <sup>70</sup>	No	100	15
	BatMag <sup>58</sup>	No	100	20
	CGCI <sup>68</sup>	No	100	8
	Orthogonally Aligned <sup>56</sup>	No	40	6
	DeltaMag <sup>62</sup>	Yes	23	–
	BigMag <sup>60</sup>	Yes	40	25
	ARMM <sup>61</sup>	Yes	80	7
Local	MiniMag <sup>71</sup>	No	20	100
	MILiMAC <sup>25</sup>	No	8	370
	MagNeed <sup>26</sup>	No	9	380
	μMAZE <sup>24</sup>	No	0.15	8000

endoscope channel<sup>25</sup>. Typically, the deployment mechanism must be a small apparatus activated/deactivated using external stimuli such as mechanical force transmitted through cables<sup>72</sup>, magnetic fields<sup>73</sup>, or temperature gradients to trigger the shape memory effect of advanced materials<sup>74,75</sup>.

For laparoscopy probes, miniaturized electromagnets can be designed in the shape of needles to facilitate their assembly<sup>26</sup>. Unlike endoscopes, electromagnet deployment and retrieval are achieved using standard pivot points, such as trocars<sup>76</sup>. The advantage of using these pivots is that the spatial configuration of the probes can be actively modified to change the workspace configuration (Fig. 2B). Furthermore, laparoscopy probes with assembled electromagnets can be attached to robotic platforms to enable object avoidance and image-guided procedures<sup>77,78</sup>.

### 2.3 Fundamentals


In order to understand the local actuation of micro-agents using miniaturized electromagnets, we present the fundamentals of actuation through mathematical formulation. We utilize three miniaturized electromagnets throughout the analysis to achieve 3D manipulation<sup>79</sup>. However, the formulation can be extended to include more electromagnets, creating an overactuated system.

In general, the magnetic field generated by an electromagnet ( $\mathbf{B}(\mathbf{p}, I)$ ) powered with an electrical current ( $I \in \mathbb{R}$ ) exerts a wrench ( $\mathbf{W} \in \mathbb{R}^6$ ) over a micro-agent with magnetic moment ( $\boldsymbol{\mu} \in \mathbb{R}^3$ ) located at a point ( $\mathbf{p} \in \mathbb{R}^3$ ). The magnetic wrench includes the magnetic force ( $\mathbf{F} \in \mathbb{R}^3$ ) and magnetic torque ( $\mathbf{T} \in \mathbb{R}^3$ ) and is defined as

$$\mathbf{W}(\mathbf{p}) = \begin{bmatrix} \mathbf{F}(\mathbf{p}) \\ \mathbf{T}(\mathbf{p}) \end{bmatrix} = \begin{bmatrix} \nabla(\boldsymbol{\mu} \cdot \mathbf{B}(\mathbf{p}, I)) \\ \boldsymbol{\mu} \times \mathbf{B}(\mathbf{p}, I) \end{bmatrix}. \quad (1)$$

Local frames ( $\{\mathcal{N}_k\}$ , for  $k = 1, 2, 3$ ) are constructed for three miniaturized electromagnets with respect to a global reference frame ( $\{\mathcal{G}\}$ ) (Fig. 3). At a point, ( ${}^{\mathcal{G}}\mathbf{p} \in \mathbb{R}^3$ ) in the global reference frame, the magnetic field ( ${}^{\mathcal{G}}\mathbf{B}({}^{\mathcal{G}}\mathbf{p})$ ) is the sum of the magnetic field generated by each electromagnet ( ${}^{\mathcal{G}}\mathbf{B}_k({}^{\mathcal{G}}\mathbf{p})$ ). The unitary magnetic field ( ${}^{\mathcal{N}_k}\boldsymbol{\beta}_k({}^{\mathcal{N}_k}\mathbf{p})$ ) is generated in the local frame of the corresponding electromagnet. The miniaturized electromagnets operate in their linear regions (i.e., the magnitude and components of the magnetic field vary linearly with the current)<sup>80</sup>. Hence, the magnetic field in local frames is defined as

$${}^{\mathcal{N}_k}\mathbf{B}_k({}^{\mathcal{N}_k}\mathbf{p}) = {}^{\mathcal{N}_k}\boldsymbol{\beta}_k({}^{\mathcal{N}_k}\mathbf{p})I_k, \quad (2)$$



**Figure 2:** Miniaturized electromagnets integrated into surgical instruments to perform local actuation of micro-agent. **(A)** Endoscope with miniaturized electromagnets deployed around a target region. **(B)** Laparoscopic instruments with integrated needle-shaped electromagnets. The trocars permit the manipulation of the instruments to span the surroundings of a target region. **(C)** Zoom in on the workspace showing magnetic micro-agents being pulled within a vasculature toward a target region.



where  $I_k \in \mathbb{R}$ , is the current through the  $k$ th electromagnet. In order to compute the vectors  ${}^{\mathcal{G}}\mathbf{B}_k({}^{\mathcal{G}}\mathbf{p})$ , we define the rotation matrices ( ${}^{\mathcal{G}}\mathbf{R}_{\mathcal{N}_k} \in SO(3)$ ). This way,  ${}^{\mathcal{N}_k}\mathbf{p} = {}^{\mathcal{N}_k}\mathbf{R}_{\mathcal{G}} {}^{\mathcal{G}}\mathbf{p} + {}^{\mathcal{N}_k}\boldsymbol{\ell}_{\mathcal{G}}$ , where  ${}^{\mathcal{N}_k}\boldsymbol{\ell}_{\mathcal{G}} \in \mathbb{R}^3$  are the distance vectors between local and global frames (Fig. 3). Hence,  ${}^{\mathcal{G}}\mathbf{B}_k({}^{\mathcal{G}}\mathbf{p})$  is computed as follows:

$${}^{\mathcal{G}}\mathbf{B}_k({}^{\mathcal{G}}\mathbf{p}) = {}^{\mathcal{G}}\mathbf{R}_{\mathcal{N}_k} {}^{\mathcal{N}_k}\boldsymbol{\beta}_k({}^{\mathcal{N}_k}\mathbf{p})I_k. \quad (3)$$

Substituting (3) into (1), we obtain the magnetic wrench ( ${}^{\mathcal{G}}\mathbf{W}({}^{\mathcal{G}}\mathbf{p})$ ) composed of force ( ${}^{\mathcal{G}}\mathbf{F}({}^{\mathcal{G}}\mathbf{p})$ ) and torque ( ${}^{\mathcal{G}}\mathbf{T}({}^{\mathcal{G}}\mathbf{p})$ ) exerted on a micro-agent located at  ${}^{\mathcal{G}}\mathbf{p}$ :

$$\begin{cases} {}^{\mathcal{G}}\mathbf{F}({}^{\mathcal{G}}\mathbf{p}) = \boldsymbol{\mu}^T \begin{bmatrix} \frac{\partial [{}^{\mathcal{G}}\boldsymbol{\beta}_1({}^{\mathcal{G}}\mathbf{p})]}{\partial x} & \dots & \frac{\partial [{}^{\mathcal{G}}\boldsymbol{\beta}_3({}^{\mathcal{G}}\mathbf{p})]}{\partial x} \\ \frac{\partial [{}^{\mathcal{G}}\boldsymbol{\beta}_1({}^{\mathcal{G}}\mathbf{p})]}{\partial y} & \dots & \frac{\partial [{}^{\mathcal{G}}\boldsymbol{\beta}_3({}^{\mathcal{G}}\mathbf{p})]}{\partial y} \\ \frac{\partial [{}^{\mathcal{G}}\boldsymbol{\beta}_1({}^{\mathcal{G}}\mathbf{p})]}{\partial z} & \dots & \frac{\partial [{}^{\mathcal{G}}\boldsymbol{\beta}_3({}^{\mathcal{G}}\mathbf{p})]}{\partial z} \end{bmatrix} \begin{bmatrix} I_1 \\ I_2 \\ I_3 \end{bmatrix} \\ {}^{\mathcal{G}}\mathbf{T}({}^{\mathcal{G}}\mathbf{p}) = \boldsymbol{\mu} \times \sum_{k=1}^3 ({}^{\mathcal{G}}\mathbf{B}_k({}^{\mathcal{G}}\mathbf{p})) \end{cases} \quad (4)$$

The presented formulation utilizes the spatial configuration of local frames ( $\{\mathcal{N}_k\}$ ) and the set of currents ( $I_k$ ) as inputs to compute equations (2)–(4) in every actuation period. Considering the characteristics of instruments integrating miniaturized electromagnets, constraints can be applied to local frames to reduce computational complexity. In the case of an endoscope integrating such electromagnets, it is reasonable to assume that the deployment mechanism will fix the electromagnets around a target region. Consequently, local frames remain approximately stationary, and the magnetic field and gradients at point ( ${}^{\mathcal{G}}\mathbf{p}$ ) are solely determined by the currents (Fig. 3A). For laparoscopy instruments integrating miniaturized electromagnets, the local frames can change their pose according to the instrument’s motion (Fig. 3B). Hence, rotation matrices ( ${}^{\mathcal{G}}\mathbf{R}_{\mathcal{N}_k}$ ) and vector distances ( ${}^{\mathcal{N}_k}\boldsymbol{\ell}_{\mathcal{G}}$ ) are updated in every actuation period. In order to ensure accuracy and robustness and account for external disturbances, various algorithm schemes for micro-agent control can be implemented for miniaturized electromagnets<sup>81</sup>. Hence, the development of clinical imaging modalities is necessary to ensure the required feedback for precise manipulation of micro-agents.


### 3 Fluorescence Imaging for Microrobotics

#### 3.1 Background

Visualization or imaging is essential in microrobotics because current sensor technology cannot be incorporated into micro-agents due to size limitations. Imaging modalities serve as external sensors to validate micro-agent functionalities and provide feedback during procedures. Recent research in the field of microrobotics underscores the significance of advanced methods in broadening the scope of microrobotics into clinically relevant scenarios<sup>82–85</sup>. Microrobotics imaging requires three main components to detect and track micro-agents under *in vivo* conditions.

- **Spatiotemporal resolution and penetration depth:** Imaging modalities in microrobotics must enable micrometer resolution, video-rate tracking, and sufficient penetration depths to achieve in-tissue visualization<sup>35</sup>. Generally, a trade-off exists between spatial resolution and penetration depth due to tissue scattering. Therefore, an optimal balance between spatiotemporal resolution and penetration depth is essential for in-tissue microrobotic imaging.
- **Molecular imaging:** Visualization modalities must ensure the detection of micro-agents and the surrounding physical environment, including tissue and biological samples<sup>38</sup>. Studying living cells, biological processes, and tissue interacting with functionalized micro-agents requires addressing molecular imaging techniques<sup>36</sup>.
- **Non-ionizing modalities:** Prolonged exposure to ionizing radiation risks side effects on patients and clinicians<sup>86</sup>. Non-ionizing modalities can enable safe procedures using micro-agents.

Current imaging modalities for microrobotics partially address all three components. Various ionizing techniques such as X-ray computerized tomography<sup>65</sup>, positron emission tomography<sup>87</sup>, fluoroscopy<sup>88</sup>, and single-photon emission computed tomography<sup>89</sup> have been proven under *ex vivo* and *in vivo* conditions. Yet, the risk of side effects from long-term exposure motivates the use of non-ionizing techniques. Among non-ionizing modalities, fluorescence imaging<sup>38</sup>, ultrasound<sup>31</sup>, magnetic resonance imaging<sup>90</sup>, photoacoustic imaging<sup>84,91</sup>, and magnetic particle imaging<sup>92</sup> provide molecular information, revealing details about the interaction between micro-agents, biological entities, and living tissue. This



**Figure 3:** Reference frames used to mathematically describe the actuation of a magnetic micro-agent located at point  $({}^G p)$ . **A** Endoscope with integrated miniaturized electromagnets. **B** Laparoscopy instruments with integrated needle-shaped electromagnets. The local frames  $\{\mathcal{N}_k\}$ , for  $k = 1, 2, 3$  are constructed with respect to the global frame  $\{\mathcal{G}\}$ . The distances between local and global frames are  ${}^N s \ell_G$ .

review focuses on fluorescence imaging modalities and their application in microrobotics. Compared to other non-ionizing imaging techniques, the relatively high spatiotemporal resolution of fluorescence imaging makes it attractive for studying biological processes in clinical settings using microrobots<sup>35</sup>. However, a major limitation of fluorescence imaging is the restricted optical penetration depth caused by tissue scattering. Advanced fluorescence imaging modalities using high order absorption such as two- and three-photon microscopy<sup>93</sup> and tuning the phase and amplitude of the light wavefront (i.e., wavefront shaping)<sup>94</sup> represent state-of-the-art techniques for overcoming the scattering limits of tissue. Nonetheless, further efforts are required to integrate these techniques into the field of microrobotics.

### 3.2 Fluorescence

Fluorescence is a phenomenon in which a substance (fluorescent dye or fluorophore) absorbs electromagnetic radiation at one wavelength and then emits light at a longer wavelength<sup>95</sup>. This shift between wavelengths is called Stokes shift and represents the difference in wavelength between the maximum excitation ( $\lambda_{ex}$ ) and maximum emission ( $\lambda_{em}$ ) of a fluorophore:

$$\text{Stokes Shift} = \lambda_{ex} - \lambda_{em}. \quad (5)$$

In fluorescence, the emission occurs promptly after excitation, often within nanoseconds to microseconds<sup>96</sup>. This property distinguishes fluorescence from other types of photoluminescence, such as phosphorescence, where the emission persists for a longer duration after excitation. Spectrum analysis using spectrofluorometers provides intensity-wavelength plots to study the excitation and emission light characteristics of fluorophores<sup>97</sup>. The fluorescence quantum yield ( $\Phi$ ) measures the efficiency of a fluorophore to emit fluorescence upon excitation:

$$\Phi = \frac{\text{Number of photons emitted}}{\text{Number of photons absorbed}}. \quad (6)$$

The value of  $\Phi$  depends on the fluorophore and solvents utilized<sup>98</sup>. Besides, the intensity of emitted fluorescence light is proportional to the number of fluorophores in the sample and quantum yield.

Fluorescence is widely utilized in various fields, including biology, chemistry, and surgery, for diagnostics and microscopy applications<sup>99,100</sup>. Due to their specificity, sensitivity, and versatility,


fluorophores are commonly employed to label biological samples, track molecular processes, and functionalize micro-agents<sup>40</sup>. In microscopy, the excitation light is focused onto the sample, and the emitted light is collected through lenses and a series of dichroic mirrors, which redirect the light for image formation through detectors (e.g., complementary metal-oxide semiconductor cameras and photomultipliers)<sup>93,101</sup>. Figure 4A depicts a fluorescent micro-agent emitting fluorescence with a wavelength of 600 nm under excitation by a continuous light source of 530 nm. The energy diagram, including ground and excited energy states, represents a fluorescence event in which a single photon at 530 nm is absorbed to generate one photon of emitted light at 600 nm. This energy representation corresponds to one-photon absorption and is the functioning principle of many fluorescence microscopes used in biology. Fluorescence imaging modalities are the gold standard for color-coded imaging using different fluorophores to label biological samples and micro-agents<sup>7,38</sup>. However, the limited optical penetration depth, scattering, and background noise reduce the capability of using conventional fluorescence microscopes to visualize micro-agent in-tissue.

### 3.3 Two-Photon Microscopy

An advanced fluorescence modality to overcome the limited penetration depth of fluorescence imaging techniques is two-photon microscopy<sup>41,102</sup>. Unlike conventional fluorescence microscopy, two-photon microscopy utilizes the simultaneous absorption of two lower-energy photons to excite fluorophores within a specimen (Fig. 4B). In two-photon microscopy, an ultrashort-pulsed laser emits near-infrared (NIR) light photons, typically within 700–1100 nanometers of wavelength. This longer wavelength light can penetrate deeper into biological tissues with minimal scattering and absorption, allowing imaging of thicker samples or even intact living organisms<sup>103</sup>.

When two photons of NIR light coincide in space and time, their combined energy excites a fluorophore to its higher energy state<sup>104</sup>. This excitation occurs only at the focal point of the microscope, where the intensity of the laser light is highest, resulting in precise spatial localization of fluorescence emission<sup>105,106</sup>. Because fluorescence is generated only at the focal point, out-of-focus fluorescence (i.e., background noise) is





**Figure 4:** Fluorescence imaging for shallow and deep tissue. **A** Conventional fluorescence imaging utilizes the principle of one-photon excitation (excitation light at 530 nm wavelength) to generate emission light at 600 nm wavelength. One-photon excitation can be used for shallow tissue as it is prone to background noise and scattering. **B** Two-photon microscopy is an advanced fluorescence imaging modality that uses the principle of two-photon excitation. As two photons are absorbed simultaneously, the excitation light is a femtosecond-pulsed laser with 1060 nm wavelength. By using near-infrared light, the light can penetrate tissue samples deeper. Besides, the two-photon excitation only occurs at the focal spot, eliminating the background noise.

minimized, improving image contrast and resolution, especially in thick samples. Two-photon microscopy offers three essential advantages over traditional fluorescence microscopy techniques, including:

- **Deep tissue imaging:** The longer wavelength of NIR light allows deeper penetration into biological specimens, enabling imaging of structures located several hundred micrometers below the tissue surface.
- **Reduced photodamage:** Because fluorescence excitation occurs only at the focal point, photodamage to the specimen outside the focal plane is minimized, making two-photon microscopy suitable for imaging living tissues over extended periods without significant damage.
- **Increased image contrast and resolution:** By minimizing out-of-focus fluorescence, two-photon microscopy provides improved image contrast and resolution, particularly in thick or densely labeled samples.
- **Fluorophore-free imaging:** The ultrashort-pulsed laser used in two-photon microscopy can trigger autofluorescence and second-harmonic generation in biological tissue samples<sup>107</sup>. This way, fluorophore-free imaging is enabled, eliminating photobleaching effects associated with standard fluorophores<sup>108</sup>.

Two-photon microscopy is widely used in neuroscience, cell biology, and other fields where high-resolution imaging of living tissues or organisms is essential<sup>109,110</sup>. In addition, various endomicroscopy probes have been developed to diagnose and study cellular processes in freely behaving animals<sup>111</sup>. In microrobotics, the principle of two-photon absorption has mainly been used to fabricate sub-micrometer resolution structures such as lenses<sup>112</sup>, sensors<sup>113</sup>, micro-agents<sup>114</sup>, and metamaterials<sup>115</sup>. Two-photon microscopy recently garners attention using benchtop microscopes to visualize micro-agents through *ex vivo* and *in vivo* rat tissue<sup>42,43</sup>.

### 3.4 Open Challenges

Although fluorescence imaging can provide information from biological samples and tissue, the optical penetration depth remains a challenge due to tissue scattering. Two-photon microscopy has demonstrated clinical applications for diagnosis and revolutionized our understanding of biological processes by enabling the visualization

of cellular dynamics and interactions with unprecedented detail<sup>93,116</sup>. However, the optical penetration depth cannot exceed hundreds of micrometers. Using three-photon absorption has opened up new avenues for imaging tissue at penetration depths on the order of 1 mm in brain tissue<sup>117,118</sup>. Increasing the order of photon absorption enables greater penetration depth due to the longer wavelength used to trigger fluorescence. However, additional challenges must be addressed, such as the limited field of view, temporal resolution, and chromatic dispersion<sup>111</sup>.

On the other hand, wavefront shaping offers an innovative approach to overcoming the scattering limits of tissue. This technique modifies the phase and amplitude of light, enabling deep focusing through scattering media up to tens of centimeters<sup>94,119–121</sup>. The convergence of high-order photon absorption and wavefront shaping can alleviate the limited penetration depth of fluorescence microscopy. However, fluorescence collection at relatively high penetration depths is an open challenge due to the absorption probability of emitted photons in scattering media. A comprehensive study of the algorithms to reconstruct images from fluorescence-emitted light in deep tissue is required<sup>45,122</sup>.

## 4 Toward Fluorescence-Guided Manipulation

The convergence of surgical instruments integrating miniaturized electromagnets and fluorescence imaging modalities can improve the precision and efficiency of targeted drug delivery procedures via functionalized micro-agents. By overcoming physiological fluid flows and scattering, fluorescence-guided manipulation of magnetic micro-agents represents a groundbreaking approach to advancing the application of microrobotics in clinically relevant scenarios. Figure 5 depicts a tissue segment containing vasculature. The micro-agents previously perfused within the vasculature are actuated and visualized through tissue using miniature electromagnets and two-photon microscopy, respectively. Our previous study demonstrates this approach using formalin-fixed rat intestinal tissue (ileal wall)<sup>42</sup>. Figure 6A shows a close-up view of the experimental setup, including a miniaturized electromagnet<sup>26</sup>. Magnetic electrospun fibers stained with a fluorophore (Coumarin 6) are fabricated as the micro-agents used in the experiments<sup>38</sup>. Micro-agents are perfused within a microfluidic channel to visualize their interactions through two-photon microscopy. For comparison, we used two

types of samples: a plain microfluidic channel (Fig. 6B) and a microfluidic channel containing a slice of rat tissue (thickness  $\approx 600 \mu\text{m}$ ) (Fig. 6C).

Using the first sample enables clear visualization of the orientation and motion of micro-agents in response to the magnetic field and gradients produced by the miniaturized electromagnet (Fig. 6B). The experiment utilizing formalin-fixed rat tissue intensifies scattering, augmenting the complexity of visualizing micro-agents. Despite the scattering challenges, our results show that two-photon microscopy facilitates the identification of morphology and tracking of micro-agents' motion while enabling continuous image acquisition.

It is worth noting that the experiments were achieved using a relatively low frame acquisition rate (1 fps), which can limit the analysis of micro-agents moving at high dynamics. Previous *in vivo* studies demonstrate that micro-agents can move at relatively low velocities ( $< 10 \mu\text{m/s}$ ) under magnetic guidance<sup>83,123</sup>. Furthermore, incorporating optical and optomechanical technologies, such as polygonal mirrors and resonant scanners coupled with optical fibers, can drastically improve the scan rate by compromising the spatial resolution<sup>116,124</sup>. On the other hand, the field of view used for image acquisition is  $135 \mu\text{m}$ , which can be challenging for imaging bigger micro-agents. Visual servoing techniques can alleviate this shortcoming by moving the field of view according to the micro-agent motion<sup>125</sup>. Furthermore, a multimodal imaging approach can complement two-photon microscopy and provide additional information on the physical surroundings<sup>126</sup>.

## 5 Discussion


Integrating miniaturized electromagnets into surgical instruments for fluorescence-guided micro-agent manipulation represents a significant advancement in microrobotics. This approach can address the limitations of current electromagnetic systems, particularly in terms of power consumption and spatial constraints. The key strength of this approach is the potential for minimally invasive procedures (e.g., targeted drug delivery and microsurgery) using functionalized micro-agents. However, integrating miniaturized electromagnets into surgical instruments requires careful design and engineering to ensure optimal performance and compatibility. Factors such as electric and thermal insulation and deployment/retrieval mechanisms must be carefully

considered to minimize the risk of complications during procedures.


We introduce fluorescence imaging modalities for magnetic micro-agent control, highlighting the micrometer resolution and the capability of obtaining feedback to study biological entities. Fluorescence imaging modalities permit color-coded image acquisition, which is appealing for visualization of micro-agents and physical surroundings without segmentation algorithms. The main drawbacks of conventional fluorescence modalities are the limited penetration depth ( $< 100 \mu\text{m}$ ) and out-of-focus fluorescence (i.e., background noise).

Two-photon microscopy alleviates the current drawbacks of conventional fluorescence microscopy. However, several challenges remain to be addressed to realize this technology's potential. One limitation is the penetration depth, which may be insufficient for specific applications when light needs to be focused through thick tissue samples ( $> 1 \text{ mm}$ ). Three-photon absorption can improve the penetration depth up to a few millimeters, but the tissue scattering limits the application for thicker samples. An alternative approach involves actively changing the phase and amplitude of the light wavefront (i.e., wavefront shaping). This technique can correct distortions caused by scattering, permitting light to focus through samples up to tens of centimeters. Yet, the triggered fluorescence risks being absorbed by the tissue, hindering its collection for image formation. This motivates the study of fluorescence acquisition in thick samples. Furthermore, the development of endomicroscopy probes integrating high-order photon absorption and wavefront shaping technology holds potential for *in situ* visualization of internal tissues (e.g., mucosa) where required penetration depths are in the order of a few millimeters.

Recognizing the potential of miniaturized electromagnets and two-photon microscopy, we present previous results for fluorescence-guided manipulation of functionalized micro-agents. The results of this synergy showcase the visualization through a rat tissue sample of  $600 \mu\text{m}$  thick. However, further improvements are needed to overcome the scattering effects, relatively low-frame acquisition, and limited field of view. Multimodal imaging strategies, novel scanning techniques, visual servoing, and graphics processing unit parallel computing can optimize the current two-photon microscopy apparatus to be used along with surgical instruments.



**Figure 5:** Illustration of fluorescence-guided manipulation of functionalized micro-agents. The magnetic micro-agents are within a vasculature transporting fluid flow. Miniaturized electromagnets are used to actuate micro-agents, while two-photon microscopy, while femtosecond-pulsed laser) penetrates tissue and vasculature to trigger emission light from fluorescent micro-agents. The magnetic field (**B**) generated by the electromagnets exerts a force (**F**) and torque (**T**) over the micro-agent.



**Figure 6:** Fluorescence-guided manipulation of micro-agents (magnetic electrospun fibers). **A** Experimental validation platform for two-photon image acquisition of micro-agents through rat tissue. The micro-agents are attached due to the electrostatic interaction. **B** Visualization of micro-agents within a microfluidic channel. The pink dashed circle indicates attachment due to electrostatic interaction. **C** Visualization of micro-agents through rat tissue. Figure adapted from [42].



### Publisher's Note

Springer Nature remains neutral with regard to jurisdictional claims in published maps and institutional affiliations.

Springer Nature or its licensor (e.g. a society or other partner) holds exclusive rights to this article under a publishing agreement with the author(s) or other rightsholder(s); author self-archiving of the accepted manuscript version of this article is solely governed by the terms of such publishing agreement and applicable law.

### Author Contributions

J.H. took the lead in writing the manuscript with contributions from S.M. The authors thoroughly reviewed the manuscript and actively participated in discussions regarding the design of the manuscript. S.M. secured the funding for the project and provided supervision throughout its course.

### Funding

This work is supported by the European Research Council (ERC) under the European Union's Horizon 2020 Research and Innovation programme under Grant #866494 (project-MAESTRO).

### Data availability

This is a review article, and no new data were generated or analyzed. All data discussed in this article are cited and accessible from the original publications referenced herein.

### Declarations

### Ethical Approval and consent to participate

Not applicable.

### Consent for publication

Not applicable.

### Conflict of interest

The Authors declare that there is no Conflict of interest.

Received: 5 May 2024 Accepted: 26 November 2024  
Published online: 10 February 2025

### References

1. Sitti M, Ceylan H, Wenqi H, Giltinan J, Turan M, Yim S, Diller E (2015) Biomedical applications of untethered mobile milli/microrobots. *Proc IEEE* 103(2):205–224
2. Nelson BJ, Kaliakatsos IK, Abbott JJ (2010) Microrobots for minimally invasive medicine. *Annu Rev Biomed Eng* 12:55–85
3. Ceylan H, Yasa IC, Kilic U, Wenqi H, Sitti M (2019) Translational prospects of untethered medical microrobots. *Progr Biomed Eng* 1(1):012002
4. Nelson BJ, Gervasoni S, Chiu PWY, Zhang L, Zemar A (2022) Magnetically actuated medical robots: an in vivo perspective. *Proc IEEE* 110(7):1028–1037
5. Keuning JD, de Vries J, Abelmanny L, Misra S (2011) Image-based magnetic control of paramagnetic microparticles in water. In: 2011 IEEE/RSJ International Conference on Intelligent Robots and Systems pp 421–426
6. Ornes S (2017) Medical microrobots have potential in surgery, therapy, imaging, and diagnostics. *Proc Natl Acad Sci* 114(47):12356–12358
7. Steager EB, Sakar MS, Magee C, Kennedy M, Cowley A, Kumar V (2013) Automated biomanipulation of single cells using magnetic microrobots. *Int J Robot Res* 32(3):346–359
8. Wang Y, Shen J, Handschuh-Wang S, Qiu M, Du S, Ben W (2022) Microrobots for targeted delivery and therapy in digestive system. *ACS Nano* 8:8
9. Abbott JJ, Diller E, Petruska AJ (2020) Magnetic methods in robotics. *Annu Rev Control Robot Auton Syst* 3(1):57–90
10. Ebrahimi N, Bi C, Cappelleri DJ, Ciuti G, Conn AT, Faivre D, Habibi N, Hošovský A, Iacovacci V, Khalil ISM et al (2021) Magnetic actuation methods in bio/soft robotics. *Adv Func Mater* 31(11):2005137
11. Wang M, Wu T, Liu R, Zhang Z, Jun L (2023) Selective and independent control of microrobots in a magnetic field: a review. *Engineering* 8:8
12. Aghakhani A, Yasa O, Wrede P, Sitti M (2020) Acoustically powered surface-slipping mobile microrobots. *Proc Natl Acad Sci* 117(7):3469–3477
13. Kaynak M, Dirix P, Sakar MS (2020) Addressable acoustic actuation of 3d printed soft robotic microsystems. *Adv Sci* 7(20):2001120
14. Mohanty S, Lin Y-H, Paul A, van den Broek MRP, Segers T, Misra S (2024) Acoustically actuated flow in microrobots powered by axisymmetric resonant bubbles. *Adv Intell Syst* 6(1):2300465
15. Bunea A-I, Glückstad J (2019) Strategies for optical trapping in biological samples: aiming at microrobotic surgeons. *Laser Photon Rev* 13(4):1800227
16. Li D, Liu C, Yang Y, Wang L, Shen Y (2020) Micro-rocket robot with all-optic actuating and tracking in blood. *Light Sci Appl* 9(1):84
17. Zhang S, Nan F, Neale Steven L (2022) Light-driven microrobots: mechanisms and applications. *Field-Driven Micro Nanorobots Biol Med* 8:91–111

18. Longchen W, Zheyang M, Chen Y, Yuanyi Z (2021) Engineering magnetic micro/nanorobots for versatile biomedical applications. *Adv Intell Syst* 3(7):2000267
19. Sitti M, Wiersma DS (2020) Pros and cons: magnetic versus optical microrobots. *Adv Mater* 32(20):1906766
20. Tiantian X, Jiangfan Y, Yan X, Choi H, Zhang L (2015) Magnetic actuation based motion control for microrobots: an overview. *Micromachines* 6(9):1346–1364
21. Heunis C, Sikorski J, Misra S (2018) Flexible instruments for endovascular interventions: improved magnetic steering, actuation, and image-guided surgical instruments. *IEEE Robot Autom Mag* 25(3):71–82
22. Chen R, Folio D, Ferreira A (2022) Analysis and comparison of electromagnetic microrobotic platforms for biomedical applications. *Appl Sci* 12(1):456
23. Matthews BD, LaVan DA, Overby DR, Karavitis J, Ingber DE (2004) Electromagnetic needles with submicron pole tip radii for nanomanipulation of biomolecules and living cells. *Appl Phys Lett* 85(14):2968–2970
24. Johnson BV, Esantsi N, Cappelleri DJ (2020) Design of the  $\mu$ MAZE platform and microrobots for independent control and micromanipulation tasks. *IEEE Robot Autom Lett* 5(4):5677–5684
25. Sikorski J, Mohanty S, Misra S (2020) MILiMAC: Flexible catheter with miniaturized electromagnets as a small-footprint system for microrobotic tasks. *IEEE Robot Autom Lett* 5(4):5260–5267
26. Huaroto JJ, Richter M, Malafaia M, Kim J, Kim C-S, Park J-O, Sikorski J, Misra S (2023) Magneed—needle-shaped electromagnets for localized actuation within compact workspaces. *IEEE Robot Autom Lett* 8(6):3908–3915
27. Richter M, Sikorski J, Makushko P, Zabala Y, Venkiteswaran VK, Makarov D, Misra S (2023) Locally addressable energy efficient actuation of magnetic soft actuator array systems. *Adv Sci* 10(24):2302077
28. Piñan Basualdo Franco N, van de Weerd RGJ, Sarthak M (2024) Tunable magnetic trap: Using passive elements to control magnetic microrobots. *IEEE Robot Autom Lett* 9(2):1788–1794
29. Lukas M, Michiel R, Pavlo M, Denys M, Sarthak M (2024) Miniaturized variable stiffness gripper locally actuated by magnetic fields. *Adv Intell Syst* 8:2400037
30. Nguyen KT, Go G, Jin Z, Darmawan BA, Yoo A, Kim S, Nan M, Lee SB, Kang B, Kim C-S et al (2021) A magnetically guided self-rolled microrobot for targeted drug delivery, real-time x-ray imaging, and microrobot retrieval. *Adv Healthcare Mater* 10(6):2001681
31. Pane S, Iacovacci V, Sinibaldi E, Menciassi A (2021) Real-time imaging and tracking of microrobots in tissues using ultrasound phase analysis. *Appl Phys Lett* 118(1):014102
32. Šuligoj F, Heunis CM, Mohanty S, Misra S (2022) Intravascular tracking of micro-agents using medical ultrasound: towards clinical applications. *IEEE Trans Biomed Eng* 69(12):3739–3747
33. Pane S, Faoro G, Sinibaldi E, Iacovacci V, Menciassi A (2022) Ultrasound acoustic phase analysis enables robotic visual-servoing of magnetic microrobots. *IEEE Trans Rob* 38(3):1571–1582
34. Pané S, Puigmartí-Luis J, Bergeles C, Chen X-Z, Pellicer E, Sort J, Počepcová V, Ferreira A, Nelson BJ (2019) Imaging technologies for biomedical micro-and nanoswimmers. *Adv Mater Technol* 4(4):1800575
35. Aziz A, Pane S, Iacovacci V, Koukourakis N, Czarske J, Menciassi A, Medina-Sánchez M, Schmidt OG (2020) Medical imaging of microrobots: toward in vivo applications. *ACS Nano* 14(9):10865–10893
36. Chen L, Lyu Y, Zhang X, Zheng L, Li Q, Ding D, Chen F, Liu Y, Li W, Zhang Y et al (2023) Molecular imaging: design mechanism and bioapplications. *Sci China Chem* 66(5):1336–1383
37. Yang L, Zhang Y, Wang Q, Chan K-F, Zhang L (2019) Automated control of magnetic spore-based microrobot using fluorescence imaging for targeted delivery with cellular resolution. *IEEE Trans Autom Sci Eng* 17(1):490–501
38. Kaya M, Stein F, Padmanaban P, Zhang Z, Rouwkema J, Khalil ISM, Misra S (2022) Visualization of micro-agents and surroundings by real-time multicolor fluorescence microscopy. *Sci Rep* 12(1):1–17
39. Li J, Li X, Luo T, Wang R, Liu C, Chen S, Li D, Yue J, Cheng S, Sun DJSR (2018) Development of a magnetic microrobot for carrying and delivering targeted cells. *Sci Robot* 3(19):eaat8829
40. Kaya M, Stein F, Rouwkema J, Khalil ISM, Misra S (2021) Serial imaging of micro-agents and cancer cell spheroids in a microfluidic channel using multicolor fluorescence microscopy. *PLoS ONE* 16(6):e0253222
41. Helmchen F, Denk W (2005) Deep tissue two-photon microscopy. *Nat Methods* 2(12):932–940
42. Huaroto JJ, Capuano L, Kaya M, Hlukhau I, Assayag F, Mohanty S, Römer G, Misra S (2023) Two-photon microscopy for microrobotics: Visualization of micro-agents below fixed tissue. *PLoS ONE* 18(8):e0289725
43. Del Campo A, Fonseca CG, Droux J, Ferry Y, Frei C, Wegener S, Weber B, El Amki M, Ahmed D (2023) Ultrasound trapping and navigation of microrobots in the mouse brain vasculature. *Nat Commun* 14(1):5889
44. Yildirim M, Sugihara H, So PTC, Sur M (2019) Functional imaging of visual cortical layers and subplate in awake mice with optimized three-photon microscopy. *Nat Commun* 10(1):177
45. Horstmeyer R, Ruan H, Yang C (2015) Guidestar-assisted wavefront-shaping methods for focusing light into biological tissue. *Nat Photon* 9(9):563–571
46. Adam G, Chowdhury S, Guix M, Benjamin VJ, Chenghao B, David C (2019) Towards functional mobile microrobotic systems. *Robotics* 8(3):69
47. Shao Y, Fahmy A, Li M, Li C, Zhao W, Sienn J (2021) Study on magnetic control systems of micro-robots. *Front Neurosci* 15:736730

48. Carpi F, Pappone C (2009) Stereotaxis niobe<sup>®</sup> magnetic navigation system for endocardial catheter ablation and gastrointestinal capsule endoscopy. *Expert Rev Med Devices* 6(5):487–498
49. Ciuti G, Valdastrì P, Menciassi A, Dario P (2010) Robotic magnetic steering and locomotion of capsule endoscope for diagnostic and surgical endoluminal procedures. *Robotica* 28(2):199–207
50. Ryan P, Diller E (2017) Magnetic actuation for full dexterity microrobotic control using rotating permanent magnets. *IEEE Trans Rob* 33(6):1398–1409
51. Kim Y, Parada GA, Liu S, Zhao X (2019) Ferromagnetic soft continuum robots. *Sci Robot* 4(33):eaax7329
52. Avaneesh R, Venezian R, Kim C-S, Park J-O, Misra S, Khalil ISM (2021) Open-loop magnetic actuation of helical robots using position-constrained rotating dipole field. In: 2021 IEEE/RSJ International Conference on Intelligent Robots and Systems (IROS), pages 8545–8550
53. Chathuranga D, Lloyd P, Chandler JH, Harris RA, Valdastrì P (2023) Assisted magnetic soft continuum robot navigation via rotating magnetic fields. *IEEE Robot Autom Lett* 9(1):183–190
54. Kummer MP, Abbott JJ, Kratochvil BE, Borer R, Sengul A, Nelson BJ (2010) Octomag: An electromagnetic system for 5-dof wireless micromanipulation. *IEEE Trans Robot* 26(6):1006–1017
55. Diller E, Giltinan J, Lum GZ, Ye Z, Sitti M (2016) Six-degree-of-freedom magnetic actuation for wireless microrobotics. *Int J Robot Res* 35(1–3):114–128
56. Niu F, Li J, Ma W, Yang J, Sun D (2017) Development of an enhanced electromagnetic actuation system with enlarged workspace. *IEEE/ASME Trans Mechatron* 22(5):2265–2276
57. Salmanipour S, Diller E (2018) Eight-degrees-of-freedom remote actuation of small magnetic mechanisms. In: 2018 IEEE international conference on robotics and automation (ICRA), pages 3608–3613. IEEE
58. Ongaro F, Pane S, Scheggi S, Misra S (2019) Design of an electromagnetic setup for independent three-dimensional control of pairs of identical and nonidentical microrobots. *IEEE Trans Robot* 35(1):174–183
59. Nguyen KT, Lee H-S, Kim J, Choi E, Park J-O, Kim C-S (2022) A composite electro-permanent magnetic actuator for microrobot manipulation. *Int J Mech Sci* 229:107516
60. Sikorski J, Dawson I, Denasi A, Hekman EEG, Misra S (2017) Introducing bigmag—a novel system for 3d magnetic actuation of flexible surgical manipulators. In: 2017 IEEE International Conference on Robotics and Automation (ICRA), pages 3594–3599. IEEE
61. Sikorski J, Heunis CM, Franco F, Misra S (2019) The ARMM system: an optimized mobile electromagnetic coil for non-linear actuation of flexible surgical instruments. *IEEE Trans Magn* 55(9):1–9
62. Yang L, Du X, Yu E, Jin D, Zhang L (2019) Deltamag: an electromagnetic manipulation system with parallel mobile coils. In: 2019 International Conference on Robotics and Automation (ICRA), pages 9814–9820
63. Chatzipirpiridis G, Ergeneman O, Pokki J, Ullrich F, Fusco S, Ortega JA, Sivaraman KM, Nelson BJ, Pané S (2015) Electroforming of implantable tubular magnetic microrobots for wireless ophthalmologic applications. *Adv Healthcare Mater* 4(2):209–214
64. Go G, Jeong S-G, Yoo A, Han J, Kang B, Kim S, Nguyen KT, Jin Z, Chang-Sei Kim Yu, Seo R et al (2020) Human adipose-derived mesenchymal stem cell-based medical microrobot system for knee cartilage regeneration in vivo. *Sci Robot* 5(38):eaay6626
65. Go G, Yoo A, Nguyen KT, Nan M, Darmawan BA, Zheng S, Kang B, Kim C-S, Bang D, Lee S, Kim K-P, Kang SS, Shim KM, Kim SE, Bang S, Kim D-H, Park J-O, Choi E (2022) Multifunctional microrobot with real-time visualization and magnetic resonance imaging for chemoembolization therapy of liver cancer. *Sci Adv* 8(46):eabq8545
66. Yue Dong L, Wang ZZ, Ji F, Chan TKF, Yang H, Chan CPL, Yang Z, Chen Z, Chang WT et al (2022) Endoscope-assisted magnetic helical micromachine delivery for biofilm eradication in tympanostomy tube. *Sci Adv* 8(40):eabq8573
67. Ben W, Fung CK, Ke Y, Qianqian W, Xianfeng X, Lidong Y, Ho K, Wang Y-XJ, Joseph JYS, Philip WYC et al (2021) Endoscopy-assisted magnetic navigation of biohybrid soft microrobots with rapid endoluminal delivery and imaging. *Sci Robot* 6(52):2813
68. Gang Eli S, Lien NB, Yehoshua S, Leslie F, Laszlo F, Bruce M, David J, Fishbein MC, Carlo G, Kim SJ (2011) Dynamically shaped magnetic fields: initial animal validation of a new remote electrophysiology catheter guidance and control system. *Circul Arrhythmia Electrophysiol* 4(5):770–777
69. Lucarini G, Mura M, Ciuti G, Rizzo R, Menciassi A (2015) Electromagnetic control system for capsule navigation: Novel concept for magnetic capsule maneuvering and preliminary study. *J Med Biol Eng* 35:428–436
70. Hoang MC, Nguyen KT, Le VH, Kim J, Choi E, Kang B, Park J-O, Kim C-S (2019) Independent electromagnetic field control for practical approach to actively locomotive wireless capsule endoscope. *IEEE Trans Syst Man Cybern Syst* 51(5):3040–3052
71. Kratochvil BE, Kummer MP, Erni S, Borer R, Frutiger DR, Schürle S, Nelson BJ (2014) Minimag: a hemispherical electromagnetic system for 5-dof wireless micromanipulation. In *Experimental Robotics: The 12th International Symposium on Experimental Robotics*, pages 317–329. Springer
72. Wang Y, Cao Q, Zhu X, Wang P (2021) A cable-driven distal end-effector mechanism for single-port robotic surgery. *Int J Comput Assist Radiol Surg* 16:301–309
73. Ze Q, Shuai W, Nishikawa J, Dai J, Sun Y, Leanza S, Zemelka C, Novelino LS, Paulino GH, Zhao RR (2022) Soft robotic origami crawler. *Sci Adv* 8(13):eabm7834

74. Mattmann M, Boehler Q, Chen XZ, Pané S, Nelson BJ (2022) Shape memory polymer variable stiffness magnetic catheters with hybrid stiffness control. In 2022 IEEE/RSJ International Conference on Intelligent Robots and Systems (IROS), pages 9589–9595. IEEE
75. Thomas TL, Bos J, Huaroto JJ, Venkiteswaran VK, Misra S (2024) A magnetically actuated variable stiffness manipulator based on deployable shape memory polymer springs. *Adv Intell Syst* 6(2):2200465
76. Neto JS, Siufi DFS, Magrina JF (2016) Trocar in conventional laparoscopic and robotic-assisted surgery as a major cause of iatrogenic trauma to the patient. *Best Pract Res Clin Obstetr Gynaecol* 35:13–19
77. Huu VV, Dang QNV, Minh TDD, Van Truong L (2021) Robotic video-assisted thoracoscopic surgery using multiport triangular trocar configuration: initial experience at a single center. *J Cardiothor Surg* 16:1–9
78. Sanchez JH, Amanhoud W, Billard A, Bouri M (2023) Enabling four-arm laparoscopic surgery by controlling two robotic assistants via haptic foot interfaces. *Int J Robot Res* 42(7):475–503
79. Petruska AJ, Nelson BJ (2015) Minimum bounds on the number of electromagnets required for remote magnetic manipulation. *IEEE Trans Robot* 31(3):714–722
80. Petruska AJ, Edelmann J, Nelson BJ (2017) Model-based calibration for magnetic manipulation. *IEEE Trans Magn* 53(7):1–6
81. Jiang J, Yang Z, Ferreira A, Zhang L (2022) Control and autonomy of microrobots: Recent progress and perspective. *Adv Intell Syst* 4(5):2100279
82. Bunea A-I, Taboryski R (2020) Recent advances in microswimmers for biomedical applications. *Micromachines* 11(12):1048
83. Wrede P, Degtyaruk O, Kalva SK, Deán-Ben XL, Bozuyuk U, Aghakhani A, Akolpoglu B, Sitti M, Razansky D (2022) Real-time 3d optoacoustic tracking of cell-sized magnetic microrobots circulating in the mouse brain vasculature. *Sci Adv* 8(19):eabm9132
84. Aziz A, Nauber R, Iglesias AS, Tang M, Ma L, Liz-Marzán LM, Schmidt OG, Medina-Sánchez M (2023) Nanomaterial-decorated micromotors for enhanced photoacoustic imaging. *J Micro Bio Robot* 19(1):37–45
85. Nauber R, Goudou SR, Goeckenjan M, Bornhäuser M, Ribeiro C, Medina-Sánchez M (2023) Medical microrobots in reproductive medicine from the bench to the clinic. *Nat Commun* 14(1):728
86. Hany Kasban MAM, El-Bendary DHS (2015) A comparative study of medical imaging techniques. *Int J Inf Sci Intell Syst* 4(2):37–58
87. Vilela D, Cossío U, Parmar J, Martínez-Villacorta AM, Gómez-Vallejo V, Llop J, Sánchez S (2018) Medical imaging for the tracking of micromotors. *ACS Nano* 12(2):1220–1227
88. Yang X, Shang W, Haojian L, Liu Y, Yang L, Tan R, Xinyu W, Shen Y (2020) An agglutinate magnetic spray transforms inanimate objects into millirobots for biomedical applications. *Sci Robot* 5(48):eabc8191
89. Iacovacci V, Blanc A, Huang H, Ricotti L, Schibli R, Mencias A, Behe M, Pané S, Nelson BJ (2019) High-resolution spect imaging of stimuli-responsive soft microrobots. *Small* 15(34):1900709
90. Martel S, Felfoul O, Mathieu J-B, Chanu A, Tamaz S, Mohammadi M, Mankiewicz M, Tabatabaei N (2009) Mri-based medical nanorobotic platform for the control of magnetic nanoparticles and flagellated bacteria for target interventions in human capillaries. *Int J Robot Res* 28(9):1169–1182
91. Zhiguang W, Li L, Yang Y, Peng H, Li Y, Yang S-Y, Wang LV, Gao W (2019) A microrobotic system guided by photoacoustic computed tomography for targeted navigation in intestines in vivo. *Sci Robot* 4(32):eaax0613
92. Saritas EU, Goodwill PW, Croft LR, Konkle JJ, Kuan L, Zheng B, Conolly SM (2013) Magnetic particle imaging (mpi) for nmr and mri researchers. *J Magn Reson* 229:116–126
93. Abdeladim L, Matho KS, Clavreul S, Mahou P, Sintès J-M, Solinas X, Arganda-Carreras I, Turney SG, Lichtman JW, Chessel A et al (2019) Multicolor multiscale brain imaging with chromatic multiphoton serial microscopy. *Nat Commun* 10(1):1–14
94. Shen Y, Liu Y, Ma C, Wang LV (2016) Focusing light through biological tissue and tissue-mimicking phantoms up to 9.6 cm in thickness with digital optical phase conjugation. *J Biomed Opt* 21(8):085001–085001
95. Lichtman JW, Conchello J-A (2005) Fluorescence microscopy. *Nat Methods* 2(12):910–919
96. Yang M, Guo X, Mou F, Guan J (2022) Lighting up micro-/nanorobots with fluorescence. *Chem Rev* 123(7):3944–3975
97. Sednev MV, Belov VN, Hell SW (2015) Fluorescent dyes with large stokes shifts for super-resolution optical microscopy of biological objects: a review. *Methods Appl Fluoresc* 3(4):042004
98. Velapoldi RA, Tønnesen HH (2004) Corrected emission spectra and quantum yields for a series of fluorescent compounds in the visible spectral region. *J Fluoresc* 14:465–472
99. Ntziachristos V (2006) Fluorescence molecular imaging. *Annu Rev Biomed Eng* 8:1–33
100. Nagaya T, Nakamura YA, Choyke PL, Kobayashi H (2017) Fluorescence-guided surgery. *Front Oncol* 7:314
101. Cang H, Yang L, Jianhua X (2024) Mosaic-picasso: accurate crosstalk removal for multiplex fluorescence imaging. *Bioinformatics* 2:784
102. Oheim M, Beaurepaire E, Chaigneau E, Mertz J, Charpak S (2001) Two-photon microscopy in brain tissue: parameters influencing the imaging depth. *J Neurosci Methods* 111(1):29–37
103. Luu P, Fraser SE, Schneider F (2024) More than double the fun with two-photon excitation microscopy. *Commun Biol* 7(1):364

104. König K (2000) Multiphoton microscopy in life sciences. *J Microsc* 200(2):83–104
105. Koenig K, Riemann I (2003) High-resolution multiphoton tomography of human skin with subcellular spatial resolution and picosecond time resolution. *J Biomed Opt* 8(3):432–439
106. Carriles R, Schafer DN, Sheetz KE, Field JJ, Cisek R, Barzda V, Sylvester AW, Squier JA (2009) Invited review article: Imaging techniques for harmonic and multiphoton absorption fluorescence microscopy. *Rev Sci Instrum* 80(8):081101
107. Zheng SW, Rachel ZCT, Yin OL, Benjamin YSG, Jirong L, Anantharaman V, Paulraj TT, Yee TH (2022) Multiphoton microscopy for the evaluation of interstitial fibrosis in extended criteria donor kidneys: a proof-of-concept study. *Clin Transplant* 8:e14717
108. Demchenko AP (2020) Photobleaching of organic fluorophores: quantitative characterization, mechanisms, protection. *Methods Appl Fluoresc* 8(2):022001
109. So PTC, Dong CY, Masters BR, Berland KM (2000) Two-photon excitation fluorescence microscopy. *Annu Rev Biomed Eng* 2(1):399–429
110. Svoboda K, Yasuda R (2006) Principles of two-photon excitation microscopy and its applications to neuroscience. *Neuron* 50(6):823–839
111. Vytautas K, Werner MP, Thomas S-R, Frédéric L, van Zandvoort Marc AMJ (2021) Two-photon endoscopy: state of the art and perspectives. *Mol Imaging Biol* 2:1–15
112. Gonzalez-Hernandez D, Varapnickas S, Bertoncini A, Liberale C, Malinauskas M (2023) Micro-optics 3d printed via multi-photon laser lithography. *Adv Opt Mater* 11(1):2201701
113. Gao H, An J, Chua CK, Bourell D, Kuo C-N, Tan DTH (2023) 3d printed optics and photonics: processes, materials and applications. *Mater Today* 69:107–132
114. Dabbagh SR, Sarabi MR, Birtek MT, Seyfi S, Sitti M, Tasoglu S (2022) 3d-printed microrobots from design to translation. *Nat Commun* 13(1):1–24
115. Kadic M, Milton GW, van Hecke M, Wegener M (2019) 3d metamaterials. *Nat Rev Phys* 1(3):198–210
116. Park H-C, Guan H, Li A, Yue Y, Li M-J, Hui L, Li X (2020) High-speed fiber-optic scanning nonlinear endomicroscopy for imaging neuron dynamics in vivo. *Opt Lett* 45(13):3605–3608
117. Horton NG, Wang K, Kobat D, Clark CG, Wise FW, Schaffer CB, Chris X (2013) In vivo three-photon microscopy of subcortical structures within an intact mouse brain. *Nat Photon* 7(3):205–209
118. Wang T, Chris X (2020) Three-photon neuronal imaging in deep mouse brain. *Optica* 7(8):947–960
119. Scharf E, Dremel J, Kuszmierz R, Czarske J (2020) Video-rate lensless endoscope with self-calibration using wavefront shaping. *Opt Lett* 45(13):3629–3632
120. Gigan S, Katz O, De Aguiar HB, Andresen ER, Aubry A, Bertolotti J, Bossy E, Bouchet D, Brake J, Brasselet S et al (2022) Roadmap on wavefront shaping and deep imaging in complex media. *J Phys: Photon* 4(4):042501
121. Ding C, Shao R, He Q, Li LS, Yang J (2024) Wavefront shaping improves the transparency of the scattering media: a review. *J Biomed Opt* 29(S1):S11507–S11507
122. Yeminy T, Katz O (2021) Guidestar-free image-guided wavefront shaping. *Sci Adv* 7(21):eabf5364
123. Zehao W, Zhang Y, Ai N, Chen H, Ge W, Qingsong X (2022) Magnetic mobile microrobots for upstream and downstream navigation in biofluids with variable flow rate. *Adv Intell Syst* 4(7):2100266
124. Boutilier RM, Park JS, Lee H (2018) High-speed two-photon laser scanning microscopy imaging of in vivo blood cells in rapid circulation at velocities of up to 1.2 millimeters per second. *Curr Opt Photon* 2(6):595–605
125. Rosa B, Erden MS, Vercauteren T, Herman B, Szewczyk J, Morel G (2012) Building large mosaics of confocal edomicroscopic images using visual servoing. *IEEE Trans Biomed Eng* 60(4):1041–1049
126. Li J, Thiele S, Kirk RW, Quirk BC, Hoogendoorn A, Chen YC, Peter K, Nicholls SJ, Verjans JW, Psaltis PJ et al (2022) 3d-printed micro lens-in-lens for in vivo multimodal microendoscopy. *Small* 18(17):2107032



**Juan J. Huaroto** received the bachelor's degree in Mechanical Engineering from the National University of Engineering, Lima, Peru. He earned his master's degree with Highest Distinction in Robotics and Automation from Université Bourgogne Franche-Comté, Besançon, France, in 2021. During the master's studies, he conducted research at the Département Automatique et Systèmes Micro-Mécatroniques (AS2M),

FEMTO-ST Institute (CNRS), Besançon, France. He subsequently received the Erasmus+ and ISITE-BFC Excellence mobility scholarships to undertake a research internship at the Transfers, Interfaces & Processes (TIPs) Laboratory, Université Libre de Bruxelles, Brussels, Belgium. He is currently pursuing a doctoral degree with the Surgical Robotics Laboratory at the University of Twente, The Netherlands. His research focuses on the development of advanced probes for actuation and imaging of microrobots.





**Sarthak Misra** received the master's degree in mechanical engineering from McGill University, Canada, and the Ph.D. degree in mechanical engineering from Johns Hopkins University, USA, in 2009. He is currently a Full Professor with the

Department of Biomechanical Engineering, University of Twente, Enschede, The Netherlands, and also affiliated with the Department of Biomaterials and Biomedical Technology, University of Groningen and University Medical Centre Groningen, The Netherlands. Prior to commencing his

Ph.D. studies, he was a Dynamics and Controls Analyst on the International Space Station Program. His research interests include surgical robotics and medical microrobotics. He was a recipient of the European Research Council Starting, Proof-of-Concept, and Consolidator Grants and the Netherlands Organization for Scientific Research VENI and VIDI Awards. He was the Co-Chair of the Robotics and Automation Society Technical Committee on Surgical Robotics and the International Federation of Automatic Control Technical Committee on Biological and Medical Systems.